# How to measure the zeta potential of individual Nanoparticles (NPs)?

To anticipate the answer: With the



#### Introduction

ZetaView®® tracks the particles individually in a Laser Scattering / Fluorescing Video Microscope. By following the Brownian motion of the particles their diffusion size is deduced¹.². As soon as an electric field is applied, the particles start an electrophoretic movement. From the unidirectional movement, the electrophoretic mobility  $\mu = v/E$  is derived, v being the velocity of the particles and E the applied electric filed. With some assumptions, sometimes philosophic, the so-called zeta potential  $\zeta$  is derived (ZP). When we speak of ZP, it is only about particles moving in a polar liquid like water. This is more than Pareto for the reason, that ZP in any other liquids like organic represents high acrobatic thinking.

The higher the ZP (positive or negative), the higher is the repulsion force between the particles and the lower the probability for a collision with subsequent aggregating. As a simple aid to memory – not always correct – stands "30 mV". Repulsion is dominant above 30 mV. For the aggregation, the Van der Waals attraction between surfaces is responsible. It starts below 30 mV and increases the nearer zeta potential approaches zero.

## Literature

All this is written in detail first by famous scientists like Smoluchowski, Einstein, Stokes and colleagues, later by Hunter, Lagally and others. An excellent description is given in the ISO documents 13099:2008, part 1 (general) and 2 (optical methods). Part 3 (organic solvents) is in preparation. ASTM\_E2834-12 is the corresponding American Standard. For literature references we recommend again the mentioned standard publication. Articles In simpler depiction you may find in our home page.

# Focus position control for the micro- electrophoresis set-up in the Particle Metrix ZetaView®®

In a 90° scattering/ fluorescing microscope like ZetaView®, the electrophoretic mobility measurement is accompanied by the electro-osmosis effect. Electro-osmotic flux of the medium is a reaction of the mobile counter- ions in the liquid to the electric field. These counter ions carry a charge opposite to the cell wall. The electro-osmosis therefore reflects the polarity and amount of the ions sitting on the surface of the wall. The migrating ions drive the whole liquid, with the liquid also the suspended particles. Electrophoresis and electro-osmosis motion superimpose each other. The separation of both effects is possible only in the so-called "stationary layers", where electro-

- 1) ISO 22412:2008 Particle Size Analysis Dynamic Light scattering (DLS)
- 2) ASTM2834-12 Standard Guide for Measurement of Particle Size Distribution of Nanomaterials in Suspension by Nanoparticle Tracking Analysis (NTA).

osmosis is zero. These layers are shown by the 2 dashed vertical lines in <u>Fig. 4 of the brochure</u> (see below).

From looking to the velocity (here zeta potential) profiles of the liquid in figure 4, it is clear, that the location of the combined focus of the laser and the microscope must be controlled over the entire cross section of the cell. Scanning from one side (wall 0) to the other side (wall 1) is a must. The velocity profile is recorded and the electrophoretic mobility (zeta potential) taken from the 2 stationary layers. Particle Metrix takes full control of it. The user has not to take care about electroosmosis influences.

If the sample is very salty like a sample in PBS, convection may





Fig. 4: Green: ZP of  $Al_2O_3$  is +50 mV, walls are cationically coated. Red: -25 mV ZP anionic polysterene, -; the walls are neutral. Blue: polystyrene -40 mV ZP, cell walls are anionic as uncoated glass usually is.


occur. To avoid this, the field is alternated at a sufficiently high repetition rate, so that the electrophoretic motion of the particles is measurable, whereas the motion of the total sample due to electro-osmosis cannot follow.

This scan helps not only in zeta potential analysis but is useful also to guaranty statistically valid size and concentration analyses by measuring concentration and size at all the 11 stops during a cell profile scan. The automatic measurement is shown below (or Video):

- 1) ISO 22412:2008 Particle Size Analysis Dynamic Light scattering (DLS)
- 2) ASTM2834-12 Standard Guide for Measurement of Particle Size Distribution of Nanomaterials in Suspension by Nanoparticle Tracking Analysis (NTA).




Programming a kinetic measurement where each profile scan is followed by a burst of 5  $\mu$ l new sample, the statistical validity is boosting. See the next slide (or Video):




- 1) ISO 22412:2008 Particle Size Analysis Dynamic Light scattering (DLS)
- 2) ASTM2834-12 Standard Guide for Measurement of Particle Size Distribution of Nanomaterials in Suspension by Nanoparticle Tracking Analysis (NTA).

#### How to measure the size and electrophoretic component of an individual particle simultaneously?

An SOP starts the zeta potential measurement at the stationary layers. The rest is mathematical separation of the uni-directional electrophoretic movement from the random Brownian motion. This is documented as a scatter plot below.



The following example shows the kinetics of zeta potential and size after adding 2 mM NaCl solution to a dispersion of 40 nm gold particles.



# **Preparing the measurement**

As the surface charge of a particle interacts strongly with the charge in its environment, a clean cell is a must. Invisible surfactants may coat the surface and change the ionic charge. During diluting the sample to the measurement concentration, the pH may change. The robust quartz cell guarantees easy cleaning. Internal pumps assist the rinsing process.

- 1) ISO 22412:2008 Particle Size Analysis Dynamic Light scattering (DLS)
- 2) ASTM2834-12 Standard Guide for Measurement of Particle Size Distribution of Nanomaterials in Suspension by Nanoparticle Tracking Analysis (NTA).

Depending on what sizes will be important to analyze, it is necessary to bring the concentration in balance with the camera settings. For known samples the SOP helps, for unknown the software contains a dilution assistant.

### Summary

For about 100 years it was not possible to measure zeta potential routinely with a micro-electrophoresis set-up like in ZetaView®. It was too complicated to set the optics to the stationary layers and to keep the focus sharp over a long time. With ZetaView® this is possible, even in combination with the Brownian motion size evaluation. As the ZetaView® was designed as a portable instrument, automation was a must. This is a strength of the instrument and guarantees repeatability and precision in concentration, size and sub-population analysis of nanoparticles, in scattering and fluorescent mode. To plot zeta potential with the new software 8.04.02 vs size tops all that.

- 1) ISO 22412:2008 Particle Size Analysis Dynamic Light scattering (DLS)
- 2) ASTM2834-12 Standard Guide for Measurement of Particle Size Distribution of Nanomaterials in Suspension by Nanoparticle Tracking Analysis (NTA).